

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1423

Name: RITTER 1  
Sterilizer ID: V2178198  
12/06/2021 10:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.6 F 0.0  
2:00 140.6 F 0.1  
4:00 146.7 F 0.3  
6:00 166.8 F 3.6  
8:00 190.0 F 7.7  
10:00 209.6 F 12.1  
12:00 223.6 F 6.3  
14:00 235.5 F 11.0  
16:00 233.3 F 7.7  
18:00 244.6 F 12.3  
20:00 244.3 F 12.6  
22:00 256.6 F 19.0  
24:00 267.7 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.4 F 28.9  
2:00 272.2 F 29.0  
3:00 271.9 F 28.5  
4:00 272.7 F 29.2

Min 271.5 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.0 F 0.0  
5:00 198.0 F 0.0  
10:00 203.8 F 0.0  
15:00 209.1 F 0.0  
20:00 208.1 F 0.0  
25:00 203.0 F 0.0  
30:00 197.6 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 25:20  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:03:45  
CYCLE COMPLETE

12/06/2021 11:45 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02