
Midmark M11 - v1.0.5
Total Cycles: 1278

Name: RITTER2
Sterilizer ID: V2313228
07/11/2022 07:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.0 F 0.0
2:00 72.4 F 0.0
4:00 77.7 F 0.0
6:00 97.2 F 0.0
8:00 128.8 F 0.1
10:00 177.0 F 0.4
12:00 198.9 F 2.0
14:00 208.4 F 5.2
16:00 219.0 F 8.4
18:00 228.7 F 12.0
20:00 230.7 F 7.8
22:00 242.2 F 12.6
24:00 239.4 F 10.6
26:00 236.6 F 9.1
28:00 250.1 F 15.6
30:00 261.0 F 21.8
32:00 271.1 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.4 F 29.3
2:00 271.5 F 28.7
3:00 272.4 F 29.4
4:00 271.6 F 28.6

Min 271.5 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 203.9 F 0.0
10:00 219.7 F 0.0
15:00 216.0 F 0.0
20:00 212.9 F 0.0
25:00 221.3 F 0.0
30:00 226.8 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:44
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:10:41
CYCLE COMPLETE

07/11/2022 08:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____