

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2438

Name: RITTER2  
Sterilizer ID: V2313228  
05/30/2023 03:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.0 F 0.0  
2:00 195.8 F 3.8  
4:00 201.5 F 9.3  
6:00 220.8 F 7.4  
8:00 232.3 F 10.9  
10:00 233.8 F 8.7  
12:00 239.3 F 10.1  
14:00 247.9 F 14.6  
16:00 260.4 F 21.6  
18:00 271.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.6 F 29.5  
2:00 271.4 F 28.8  
3:00 272.8 F 29.6  
4:00 271.9 F 29.0  
  
Min 271.4 F 28.8  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.7 F 0.0  
5:00 227.9 F 0.0  
10:00 227.7 F 0.0  
15:00 225.0 F 0.0  
20:00 216.7 F 0.0  
25:00 216.7 F 0.0  
30:00 218.4 F 0.0  
Drying Complete

FILLING: 2:23  
HEATING: 18:35  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:12  
CYCLE COMPLETE

05/30/2023 03:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_