
Midmark M11 - v1.0.0
Total Cycles: 1190

Name: RITTER 1
Sterilizer ID: V2178198
09/21/2021 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.4 F 0.0
2:00 193.0 F 3.2
4:00 193.0 F 8.2
6:00 211.6 F 10.8
8:00 220.5 F 9.5
10:00 236.2 F 8.7
12:00 237.9 F 12.7
14:00 241.5 F 10.8
16:00 254.2 F 17.3
18:00 265.4 F 23.9
20:00 270.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.0
1:00 273.1 F 29.3
2:00 272.8 F 29.5
3:00 271.5 F 29.1
4:00 272.4 F 29.0

Min 271.4 F 28.9
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 239.3 F 0.0
10:00 226.6 F 0.0
15:00 227.0 F 0.0
20:00 229.9 F 0.0
25:00 230.0 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 20:38
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

09/21/2021 10:44 AM
ACCEPTED: X
REJECTED:
Operator: U01