
Midmark M11 - v1.0.5
Total Cycles: 1629

Name: RITTER2
Sterilizer ID: V2313228
10/20/2022 09:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.6 F 0.0
2:00 182.6 F 3.1
4:00 200.6 F 8.4
6:00 231.7 F 9.5
8:00 236.6 F 10.8
10:00 235.0 F 8.7
12:00 239.3 F 9.9
14:00 248.5 F 14.8
16:00 261.0 F 21.9
18:00 272.4 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.5
1:00 272.4 F 29.7
2:00 272.0 F 29.8
3:00 272.3 F 29.7
4:00 272.1 F 29.5

Min 271.6 F 29.4
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 231.4 F 0.0
10:00 224.5 F 0.0
15:00 230.5 F 0.0
20:00 227.1 F 0.0
25:00 233.4 F 0.0
30:00 202.4 F 0.0
Drying Complete

FILLING: 2:46
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

10/20/2022 10:25 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____