
Midmark M11 - v1.0.5
Total Cycles: 3869

Name: RITTER2
Sterilizer ID: V2313228
08/16/2024 03:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.6 F 0.0
2:00 172.1 F 2.5
4:00 175.3 F 6.5
6:00 200.9 F 11.2
8:00 216.5 F 5.7
10:00 230.1 F 10.9
12:00 231.0 F 8.2
14:00 239.9 F 11.3
16:00 244.5 F 13.3
18:00 257.0 F 19.8
20:00 267.7 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.6 F 29.9
2:00 271.6 F 29.2
3:00 271.5 F 29.7
4:00 273.0 F 30.1

Min 270.9 F 29.0
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 207.1 F 0.0
10:00 208.3 F 0.0
15:00 208.4 F 0.0
20:00 208.4 F 0.0
25:00 201.1 F 0.0
30:00 199.9 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 21:24
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:00:44
CYCLE COMPLETE

08/16/2024 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____