
Midmark M11 - v1.0.0
Total Cycles: 4392

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2024 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.8 F 0.0
2:00 140.6 F 1.8
4:00 159.4 F 4.8
6:00 181.7 F 8.4
8:00 200.3 F 13.2
10:00 214.9 F 5.0
12:00 225.7 F 9.0
14:00 236.6 F 13.4
16:00 237.4 F 10.0
18:00 238.0 F 9.2
20:00 247.4 F 13.9
22:00 258.7 F 20.0
24:00 268.9 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.2
3:00 272.5 F 29.0
4:00 271.8 F 28.4

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 220.9 F 0.0
10:00 223.0 F 0.0
15:00 220.3 F 0.0
20:00 219.6 F 0.0
25:00 220.4 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 25:09
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:04:01
CYCLE COMPLETE

03/21/2024 02:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____