

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001619  
16:55 20/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 2:56  
STERILIZING 4:00  
136.5 C 226kPa 4:00  
136.4 C 226kPa 4:30  
136.2 C 224kPa 5:00  
136.0 C 222kPa 5:30  
136.4 C 224kPa 6:00  
136.4 C 224kPa 6:30  
136.5 C 225kPa 7:00  
136.3 C 225kPa 7:30  
136.3 C 225kPa 7:30  
VENTING 7:31  
AIR DRYING 8:25  
CYCLE COMPLETE 8:39  
-----

Digital Signature #  
3A68B7499C3B8A55  
-----