
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001065
15:02 06/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:22
STERILIZING 3:22
136.0 C 220kPa 3:22
135.8 C 220kPa 3:52
136.3 C 225kPa 4:22
136.4 C 225kPa 4:52
136.3 C 224kPa 5:22
136.3 C 225kPa 5:52
136.1 C 221kPa 6:22
136.2 C 223kPa 6:52
136.2 C 223kPa 6:52
VENTING 6:53
AIR DRYING 7:42
CYCLE COMPLETE 8:52

Digital Signature #
52D40A4B440F07F8
