

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000221  
14:25 22/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:01  
PRESSURIZING 4:02  
STERILIZING 5:18  
136.0 C 223kPa 5:18  
136.1 C 232kPa 5:48  
136.3 C 233kPa 6:18  
136.3 C 232kPa 6:48  
136.4 C 233kPa 7:18  
136.3 C 230kPa 7:48  
136.4 C 233kPa 8:18  
136.2 C 232kPa 8:48  
136.2 C 232kPa 8:48  
VENTING 8:48  
AIR DRYING 9:51  
CYCLE COMPLETE 23:22  
-----

Digital Signature #  
FA34DD7567089DD8  
-----