
Midmark M11 - v1.0.5
Total Cycles: 2207

Name: RITTER2
Sterilizer ID: V2313228
04/03/2023 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.5 F 0.0
2:00 154.2 F 2.4
4:00 166.9 F 5.7
6:00 187.0 F 10.4
8:00 213.8 F 5.0
10:00 227.5 F 9.7
12:00 232.4 F 8.9
14:00 241.1 F 12.7
16:00 241.0 F 11.5
18:00 254.2 F 18.2
20:00 265.8 F 25.3
22:00 272.4 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.7
1:00 272.6 F 29.9
2:00 272.9 F 30.4
3:00 272.0 F 29.6
4:00 272.8 F 30.2

Min 271.9 F 29.6
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 224.7 F 0.0
10:00 231.3 F 0.0
15:00 229.9 F 0.0
20:00 228.9 F 0.0
25:00 228.7 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:26
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:42
CYCLE COMPLETE

04/03/2023 12:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____