

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000135  
17:18 08/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:35  
PRESSURIZING 4:35  
STERILIZING 5:44  
136.0 C 220kPa 5:44  
136.4 C 223kPa 6:14  
136.3 C 224kPa 6:44  
136.2 C 224kPa 7:14  
136.3 C 224kPa 7:44  
136.4 C 224kPa 8:14  
136.4 C 223kPa 8:44  
136.3 C 225kPa 9:14  
136.3 C 225kPa 9:14  
VENTING 9:14  
AIR DRYING 10:15  
CYCLE COMPLETE 10:43  
-----

Digital Signature #  
63E6BEA987341EC7  
-----