

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4508

Name: RITTER2  
Sterilizer ID: V2313228  
04/11/2025 04:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.2 F 0.0  
2:00 116.8 F 1.0  
4:00 119.2 F 3.1  
6:00 164.7 F 6.2  
8:00 173.4 F 10.1  
10:00 205.7 F 7.2  
12:00 219.9 F 6.3  
14:00 231.8 F 10.2  
16:00 231.9 F 8.1  
18:00 239.9 F 11.1  
20:00 236.1 F 8.9  
22:00 248.7 F 14.9  
24:00 259.2 F 20.8  
26:00 268.7 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.3 F 29.4  
2:00 271.5 F 29.6  
3:00 272.4 F 29.6  
4:00 271.9 F 29.4

Min 270.8 F 28.7  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.5 F 0.0  
5:00 243.1 F 0.0  
10:00 228.1 F 0.0  
15:00 222.9 F 0.0  
20:00 223.6 F 0.0  
25:00 188.7 F 0.0  
30:00 166.3 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 26:57  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:05:51  
CYCLE COMPLETE

04/11/2025 06:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_