

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4273

Name: RITTER 1  
Sterilizer ID: V2178198  
02/06/2024 09:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.2 F 0.0  
2:00 86.9 F 0.3  
4:00 88.4 F 1.2  
6:00 120.3 F 3.3  
8:00 142.5 F 6.1  
10:00 167.1 F 9.6  
12:00 187.5 F 13.5  
14:00 213.0 F 3.7  
16:00 218.8 F 6.9  
18:00 228.8 F 10.3  
20:00 232.3 F 8.8  
22:00 238.3 F 10.1  
24:00 237.6 F 8.9  
26:00 246.9 F 13.3  
28:00 258.0 F 19.3  
30:00 268.1 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.4 F 28.3  
2:00 271.9 F 28.4  
3:00 271.8 F 28.1  
4:00 272.4 F 28.6

Min 271.4 F 27.8  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.9 F 0.0  
5:00 200.9 F 0.0  
10:00 198.8 F 0.0  
15:00 204.5 F 0.0  
20:00 240.0 F 0.0  
25:00 225.7 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:18  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:09:24  
CYCLE COMPLETE

02/06/2024 10:24 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_