
Midmark M11 - v1.0.5
Total Cycles: 982

Name: RITTER2
Sterilizer ID: V2313228
04/13/2022 10:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.4 F 0.0
2:00 200.1 F 1.1
4:00 203.4 F 6.9
6:00 221.0 F 12.5
8:00 230.0 F 9.0
10:00 239.4 F 9.9
12:00 244.5 F 12.7
14:00 246.0 F 13.6
16:00 259.3 F 20.9
18:00 271.1 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.6 F 29.3
2:00 272.0 F 29.4
3:00 272.4 F 29.2
4:00 272.0 F 29.4

Min 271.6 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 212.5 F 0.0
10:00 224.6 F 0.0
15:00 226.8 F 0.0
20:00 229.6 F 0.0
25:00 229.8 F 0.0
30:00 226.3 F 0.0

Drying Complete

FILLING: 1:33
HEATING: 18:41
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:24
CYCLE COMPLETE

04/13/2022 11:15 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____