
Midmark M11 - v1.0.5
Total Cycles: 784

Name: RITTER2
Sterilizer ID: V2313228
02/10/2022 03:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.3 F 0.0
2:00 169.3 F 3.0
4:00 189.9 F 8.1
6:00 214.1 F 10.9
8:00 227.6 F 8.7
10:00 236.7 F 10.8
12:00 241.8 F 12.4
14:00 241.9 F 11.7
16:00 255.4 F 18.8
18:00 267.3 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.7
2:00 271.4 F 29.0
3:00 272.5 F 29.9
4:00 272.1 F 29.3

Min 271.4 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 167.4 F 0.0
5:00 220.8 F 0.0
10:00 198.4 F 0.0
15:00 204.5 F 0.0
20:00 171.4 F 0.0
25:00 176.9 F 0.0
30:00 178.4 F 0.0
Drying Complete

FILLING: 3:05
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:40
CYCLE COMPLETE

02/10/2022 04:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____