
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002715
12:40 18/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:03
STERILIZING 3:59
136.4 C 224kPa 3:59
136.4 C 225kPa 4:29
136.0 C 220kPa 4:59
136.4 C 224kPa 5:29
136.3 C 223kPa 5:59
136.4 C 220kPa 6:29
136.4 C 222kPa 6:59
135.9 C 218kPa 7:29
135.9 C 218kPa 7:29
VENTING 7:29
AIR DRYING 8:23
CYCLE COMPLETE 8:27

Digital Signature #
E20BF9F1B554C96A
