
Midmark M11 - v1.0.5
Total Cycles: 198

Name: RITTER2
Sterilizer ID: V2313228
07/13/2021 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.6 F 0.0
2:00 118.9 F 0.8
4:00 131.1 F 2.9
6:00 156.4 F 6.5
8:00 181.7 F 11.1
10:00 211.0 F 4.9
12:00 220.9 F 9.6
14:00 229.8 F 9.5
16:00 238.0 F 11.5
18:00 237.3 F 9.8
20:00 250.3 F 16.1
22:00 261.3 F 22.5
24:00 271.9 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.4 F 29.8
2:00 271.3 F 29.2
3:00 272.5 F 29.9
4:00 271.6 F 29.2

Min 271.2 F 29.1
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 212.2 F 0.0
10:00 222.2 F 0.0
15:00 226.5 F 0.0
20:00 227.7 F 0.0
25:00 222.2 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 24:32
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:02:38
CYCLE COMPLETE

07/13/2021 09:19 AM
ACCEPTED: X
REJECTED:
Operator: U01