
Midmark M11 - v1.0.5
Total Cycles: 1929

Name: RITTER2
Sterilizer ID: V2313228
01/12/2023 01:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 175.2 F 3.9
4:00 193.0 F 8.6
6:00 216.3 F 8.1
8:00 225.0 F 9.0
10:00 233.9 F 9.3
12:00 242.2 F 13.0
14:00 243.0 F 12.4
16:00 256.5 F 19.6
18:00 268.9 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.5 F 30.1
2:00 272.2 F 29.5
3:00 271.2 F 29.1
4:00 271.9 F 30.0

Min 271.2 F 29.1
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 228.4 F 0.0
10:00 228.1 F 0.0
15:00 229.5 F 0.0
20:00 228.5 F 0.0
25:00 229.7 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:37
CYCLE COMPLETE

01/12/2023 02:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____