
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002361
9:36 06/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:59
STERILIZING 3:00
136.0 C 221kPa 3:00
136.4 C 225kPa 3:30
136.2 C 224kPa 4:00
136.3 C 224kPa 4:30
136.3 C 223kPa 5:00
136.2 C 223kPa 5:30
136.3 C 223kPa 6:00
136.4 C 225kPa 6:30
136.4 C 225kPa 6:30
VENTING 6:30
AIR DRYING 7:24
CYCLE COMPLETE 18:30

Digital Signature #
D2785EE48231B982
