
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000160
12:37 15/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:12
CYCLE ABORTED 5:22
CYCLE FAULT 3
NOT STERILE
END TIME 5:22

Digital Signature #
0B6991BDB7FABFC7
