
Midmark M11 - v1.0.0
Total Cycles: 3107

Name: RITTER 1
Sterilizer ID: V2178198
04/13/2023 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.6 F 0.0
2:00 165.2 F 0.9
4:00 192.4 F 4.6
6:00 205.7 F 9.0
8:00 226.9 F 7.0
10:00 227.6 F 9.1
12:00 237.2 F 10.7
14:00 240.5 F 11.3
16:00 237.1 F 9.3
18:00 250.9 F 15.8
20:00 262.1 F 22.3
22:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.0 F 29.1
2:00 272.1 F 28.9
3:00 271.4 F 28.6
4:00 272.5 F 29.3

Min 271.3 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 220.3 F 0.0
10:00 219.3 F 0.0
15:00 230.1 F 0.0
20:00 228.4 F 0.0
25:00 225.5 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:22
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:27
CYCLE COMPLETE

04/13/2023 01:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____