
Midmark M11 - v1.0.5
Total Cycles: 1576

Name: RITTER2
Sterilizer ID: V2313228
10/06/2022 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.2 F 0.0
2:00 150.8 F 0.0
4:00 168.5 F 0.6
6:00 195.4 F 5.3
8:00 212.2 F 10.7
10:00 223.5 F 6.6
12:00 236.9 F 12.1
14:00 236.8 F 10.1
16:00 236.6 F 8.7
18:00 250.1 F 15.9
20:00 262.1 F 22.7
22:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.8 F 29.7
2:00 271.8 F 29.0
3:00 272.2 F 29.7
4:00 272.7 F 29.6

Min 271.5 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 204.2 F 0.0
10:00 216.7 F 0.0
15:00 221.6 F 0.0
20:00 223.7 F 0.0
25:00 224.7 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

10/06/2022 02:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____