
Midmark M11 - v1.0.0
Total Cycles: 4289

Name: RITTER 1
Sterilizer ID: V2178198
02/09/2024 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.5 F 0.0
2:00 148.5 F 1.8
4:00 150.8 F 4.7
6:00 187.5 F 8.0
8:00 197.2 F 12.3
10:00 213.4 F 4.6
12:00 221.9 F 8.7
14:00 232.8 F 13.2
16:00 235.4 F 9.7
18:00 237.9 F 9.5
20:00 246.4 F 13.7
22:00 258.1 F 19.9
24:00 268.5 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.2
2:00 271.5 F 28.5
3:00 272.1 F 29.2
4:00 272.1 F 29.0

Min 271.2 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 227.0 F 0.0
10:00 230.9 F 0.0
15:00 229.1 F 0.0
20:00 229.3 F 0.0
25:00 229.1 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:16
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:60
CYCLE COMPLETE

02/09/2024 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____