
Midmark M11 - v1.0.0
Total Cycles: 3504

Name: RITTER 1
Sterilizer ID: V2178198
07/10/2023 09:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.1 F 0.0
2:00 184.5 F 0.3
4:00 190.1 F 3.7
6:00 211.6 F 8.7
8:00 227.8 F 11.4
10:00 231.6 F 9.3
12:00 237.5 F 10.1
14:00 243.0 F 12.4
16:00 241.2 F 10.9
18:00 254.3 F 17.6
20:00 265.5 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 28.7
2:00 272.0 F 28.9
3:00 272.0 F 28.5
4:00 272.5 F 29.2

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 218.9 F 0.0
10:00 226.4 F 0.0
15:00 231.0 F 0.0
20:00 226.4 F 0.0
25:00 226.5 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 21:44
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:33
CYCLE COMPLETE

07/10/2023 10:34 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____