
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002583
9:25 21/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 2:00
STERILIZING 3:11
136.4 C 224kPa 3:11
136.2 C 222kPa 3:41
136.4 C 225kPa 4:11
136.1 C 223kPa 4:41
136.3 C 224kPa 5:11
136.4 C 226kPa 5:41
136.4 C 225kPa 6:11
136.1 C 222kPa 6:41
136.1 C 222kPa 6:41
VENTING 6:41
AIR DRYING 7:30
CYCLE COMPLETE 20:41

Digital Signature #
FBE74B6263EDB0C8
