
Midmark M11 - v1.0.0
Total Cycles: 2464

Name: RITTER 1
Sterilizer ID: V2178198
10/20/2022 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.6 F 0.0
2:00 192.4 F 4.7
4:00 199.6 F 10.1
6:00 219.2 F 5.3
8:00 232.1 F 10.6
10:00 233.6 F 7.6
12:00 242.9 F 11.4
14:00 245.5 F 13.3
16:00 258.5 F 20.3
18:00 270.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.4
2:00 271.7 F 28.6
3:00 272.4 F 29.4
4:00 272.5 F 29.2

Min 271.5 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 231.9 F 0.0
10:00 228.8 F 0.0
15:00 228.6 F 0.0
20:00 226.7 F 0.0
25:00 228.0 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 18:53
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

10/20/2022 12:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____