
Midmark M11 - v1.0.5
Total Cycles: 2228

Name: RITTER2
Sterilizer ID: V2313228
04/11/2023 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.9 F 0.0
2:00 150.0 F 2.2
4:00 163.8 F 5.0
6:00 186.5 F 9.3
8:00 222.7 F 7.4
10:00 222.9 F 8.7
12:00 232.8 F 11.2
14:00 237.8 F 11.2
16:00 236.4 F 9.6
18:00 250.5 F 16.2
20:00 262.5 F 23.2
22:00 272.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.2
1:00 272.4 F 30.1
2:00 272.1 F 29.6
3:00 271.1 F 29.0
4:00 272.6 F 30.5

Min 270.9 F 29.0
Max 273.2 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 229.1 F 0.0
10:00 231.3 F 0.0
15:00 223.0 F 0.0
20:00 218.7 F 0.0
25:00 227.2 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:28
CYCLE COMPLETE

04/11/2023 09:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____