
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000926
11:28 11/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:33
STERILIZING 3:42
136.0 C 229kPa 3:42
136.2 C 230kPa 4:12
136.2 C 231kPa 4:42
136.1 C 230kPa 5:12
136.2 C 232kPa 5:42
136.4 C 232kPa 6:12
136.3 C 231kPa 6:42
136.4 C 232kPa 7:12
136.4 C 232kPa 7:12
VENTING 7:12
AIR DRYING 8:01
CYCLE COMPLETE 23:01

Digital Signature #
DE72110AFF364CBF
