

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3523

Name: RITTER 1  
Sterilizer ID: V2178198  
07/13/2023 10:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.2 F 0.0  
2:00 188.7 F 2.3  
4:00 189.3 F 6.1  
6:00 212.5 F 10.9  
8:00 223.2 F 6.1  
10:00 233.5 F 11.4  
12:00 234.1 F 8.5  
14:00 240.1 F 10.5  
16:00 245.9 F 13.4  
18:00 258.0 F 19.9  
20:00 269.1 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.7  
1:00 272.6 F 29.4  
2:00 271.6 F 28.6  
3:00 272.6 F 29.5  
4:00 272.2 F 29.0  
  
Min 271.5 F 28.6  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 214.3 F 0.0  
10:00 236.2 F 0.0  
15:00 228.8 F 0.0  
20:00 225.5 F 0.0  
25:00 223.5 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 21:06  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:48  
CYCLE COMPLETE

07/13/2023 11:35 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_