
Midmark M11 - v1.0.0
Total Cycles: 946

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2021 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.4 F 0.0
2:00 191.2 F 2.5
4:00 192.8 F 6.7
6:00 207.2 F 11.9
8:00 226.8 F 7.1
10:00 235.4 F 12.2
12:00 236.0 F 9.5
14:00 239.6 F 9.4
16:00 247.1 F 13.8
18:00 258.6 F 20.1
20:00 269.0 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.6 F 29.0
2:00 272.4 F 29.2
3:00 272.1 F 29.2
4:00 271.6 F 28.7

Min 271.6 F 28.7
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 236.9 F 0.0
10:00 227.7 F 0.0
15:00 227.0 F 0.0
20:00 217.6 F 0.0
25:00 212.9 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 21:41
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

06/28/2021 11:14 AM
ACCEPTED: X
REJECTED:
Operator: U01