

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2599

Name: RITTER 1  
Sterilizer ID: V2178198  
11/22/2022 05:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.7 F 0.0  
2:00 163.5 F 2.9  
4:00 175.6 F 7.4  
6:00 198.5 F 12.6  
8:00 214.1 F 6.6  
10:00 228.0 F 12.1  
12:00 235.3 F 9.1  
14:00 238.3 F 9.4  
16:00 247.0 F 13.9  
18:00 259.1 F 20.6  
20:00 270.1 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.2  
1:00 272.5 F 29.4  
2:00 272.6 F 29.7  
3:00 271.8 F 29.3  
4:00 272.4 F 29.5

Min 271.8 F 29.1  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.3 F 0.0  
5:00 223.4 F 0.0  
10:00 230.8 F 0.0  
15:00 221.7 F 0.0  
20:00 224.0 F 0.0  
25:00 230.3 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 21:36  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:00:01  
CYCLE COMPLETE

11/22/2022 06:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_