
Midmark M11 - v1.0.5
Total Cycles: 3285

Name: RITTER2
Sterilizer ID: V2313228
01/10/2024 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.9 F 0.0
2:00 150.8 F 1.8
4:00 170.1 F 5.3
6:00 195.2 F 9.4
8:00 210.8 F 9.1
10:00 226.0 F 8.6
12:00 237.4 F 13.6
14:00 238.4 F 11.4
16:00 237.7 F 10.3
18:00 251.2 F 16.9
20:00 262.7 F 23.6
22:00 272.0 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.5
1:00 271.8 F 29.6
2:00 272.3 F 30.1
3:00 271.1 F 29.6
4:00 272.2 F 29.8

Min 271.1 F 29.4
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 225.3 F 0.0
10:00 222.8 F 0.0
15:00 214.0 F 0.0
20:00 208.2 F 0.0
25:00 206.6 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:01:16
CYCLE COMPLETE

01/10/2024 04:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____