
Midmark M11 - v1.0.0
Total Cycles: 3889

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2023 11:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.7 F 0.0
2:00 190.6 F 3.7
4:00 205.5 F 8.6
6:00 224.7 F 7.6
8:00 232.1 F 9.8
10:00 237.6 F 8.3
12:00 244.7 F 12.7
14:00 245.2 F 12.8
16:00 258.3 F 19.9
18:00 270.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 28.9
2:00 271.5 F 28.6
3:00 272.5 F 28.9
4:00 271.4 F 28.3

Min 271.3 F 28.2
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 212.3 F 0.0
10:00 237.2 F 0.0
15:00 231.5 F 0.0
20:00 218.3 F 0.0
25:00 212.4 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 18:51
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

10/11/2023 12:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____