
Midmark M11 - v1.0.0
Total Cycles: 3591

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2023 01:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.8 F 0.0
2:00 120.9 F 0.0
4:00 134.8 F 1.2
6:00 171.6 F 3.6
8:00 192.6 F 6.5
10:00 212.8 F 10.2
12:00 224.0 F 8.3
14:00 225.9 F 7.7
16:00 232.9 F 12.1
18:00 234.0 F 8.6
20:00 243.2 F 12.2
22:00 243.6 F 12.1
24:00 255.1 F 18.0
26:00 265.2 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.3 F 28.8
2:00 271.2 F 28.6
3:00 272.4 F 29.1
4:00 272.5 F 29.3

Min 271.0 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 211.2 F 0.0
10:00 230.3 F 0.0
15:00 228.1 F 0.0
20:00 225.0 F 0.0
25:00 222.0 F 0.0
30:00 211.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 27:55
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:06:17
CYCLE COMPLETE

07/27/2023 02:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____