
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000642
14:46 15/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:09
STERILIZING 3:13
136.4 C 224kPa 3:13
136.1 C 222kPa 3:43
136.2 C 220kPa 4:13
136.4 C 225kPa 4:43
136.1 C 220kPa 5:13
136.3 C 221kPa 5:43
136.2 C 222kPa 6:13
136.2 C 218kPa 6:43
136.2 C 218kPa 6:43
VENTING 6:43
AIR DRYING 7:36
CYCLE COMPLETE 12:16

Digital Signature #
7EFD16856504E43D
