
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002057
9:13 19/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:44
STERILIZING 3:48
136.4 C 224kPa 3:48
136.4 C 225kPa 4:18
136.2 C 223kPa 4:48
136.2 C 223kPa 5:18
136.3 C 224kPa 5:48
136.4 C 224kPa 6:18
136.4 C 223kPa 6:48
136.2 C 223kPa 7:18
136.2 C 223kPa 7:18
VENTING 7:19
AIR DRYING 8:08
CYCLE COMPLETE 8:13

Digital Signature #
D500133FA57A6DF7
