
Midmark M11 - v1.0.0
Total Cycles: 1878

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2022 09:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.4 F 0.0
2:00 203.7 F 3.5
4:00 207.6 F 8.2
6:00 219.5 F 8.4
8:00 230.9 F 9.1
10:00 240.7 F 9.7
12:00 242.4 F 12.3
14:00 241.3 F 10.9
16:00 254.8 F 17.8
18:00 266.1 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 29.0
2:00 271.5 F 28.4
3:00 272.7 F 29.0
4:00 271.6 F 28.3

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 215.0 F 0.0
10:00 233.3 F 0.0
15:00 229.9 F 0.0
20:00 225.6 F 0.0
25:00 225.2 F 0.0
30:00 224.7 F 0.0

Drying Complete

FILLING: 1:30
HEATING: 19:37
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

04/27/2022 10:31 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____