
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001505
9:05 10/15/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:24
PRESSURIZING 7:22
STERILIZING 8:59
136.0 C 220kPa 8:59
136.4 C 223kPa 9:29
136.4 C 218kPa 9:59
136.4 C 223kPa 10:29
136.2 C 221kPa 10:59
136.4 C 221kPa 11:29
136.4 C 222kPa 11:59
136.5 C 223kPa 12:29
VENTING 12:29
AIR DRYING 13:29
CYCLE COMPLETE 43:30

Digital Signature #
D15760E91E38A67F
