

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000884  
16:51 26/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:42  
PRESSURIZING 3:44  
STERILIZING 5:19  
136.5 C 233kPa 5:19  
136.3 C 232kPa 5:49  
136.4 C 233kPa 6:19  
136.3 C 232kPa 6:49  
136.4 C 233kPa 7:19  
136.2 C 231kPa 7:49  
136.4 C 233kPa 8:19  
136.3 C 231kPa 8:49  
136.3 C 231kPa 8:49  
VENTING 8:50  
AIR DRYING 9:45  
CYCLE COMPLETE 16:30

-----  
Digital Signature #  
716034FDD5C3730C  
-----