
Midmark M11 - v1.0.0
Total Cycles: 2154

Name: RITTER 1
Sterilizer ID: V2178198
07/20/2022 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.8 F 0.0
2:00 167.7 F 0.2
4:00 175.7 F 3.8
6:00 190.6 F 7.9
8:00 208.6 F 12.7
10:00 222.5 F 6.3
12:00 233.4 F 11.2
14:00 231.5 F 7.4
16:00 243.4 F 13.2
18:00 241.6 F 11.3
20:00 253.6 F 17.4
22:00 264.2 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 271.6 F 28.8
2:00 272.6 F 29.3
3:00 271.9 F 28.7
4:00 271.5 F 28.6

Min 271.2 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 222.7 F 0.0
10:00 229.7 F 0.0
15:00 233.5 F 0.0
20:00 222.7 F 0.0
25:00 235.2 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 23:58
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:56
CYCLE COMPLETE

07/20/2022 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____