
Midmark M11 - v1.0.0
Total Cycles: 1675

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2022 07:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.1 F 0.0
2:00 74.2 F 0.0
4:00 95.8 F 0.3
6:00 119.9 F 1.6
8:00 151.1 F 3.2
10:00 175.0 F 5.7
12:00 194.6 F 9.2
14:00 210.4 F 13.3
16:00 218.9 F 5.2
18:00 229.9 F 9.2
20:00 240.3 F 13.5
22:00 238.9 F 10.3
24:00 236.1 F 8.0
26:00 249.9 F 14.9
28:00 261.8 F 21.7
30:00 272.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.8 F 28.7
2:00 271.6 F 27.9
3:00 272.9 F 28.8
4:00 272.0 F 28.0

Min 271.6 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.8 F 0.0
5:00 203.2 F 0.0
10:00 209.9 F 0.0
15:00 208.4 F 0.0
20:00 207.0 F 0.0
25:00 206.9 F 0.0
30:00 206.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:26
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 01:08:11
CYCLE COMPLETE

02/28/2022 08:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____