

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4140

Name: RITTER2  
Sterilizer ID: V2313228  
11/20/2024 11:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.8 F 0.0  
2:00 160.4 F 1.5  
4:00 172.2 F 4.6  
6:00 194.5 F 8.6  
8:00 212.4 F 12.8  
10:00 218.5 F 6.9  
12:00 231.7 F 11.4  
14:00 232.2 F 8.0  
16:00 244.0 F 13.5  
18:00 243.0 F 12.1  
20:00 254.2 F 18.1  
22:00 264.7 F 24.2  
24:00 271.8 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.5  
1:00 272.5 F 29.6  
2:00 272.8 F 29.9  
3:00 271.5 F 29.7  
4:00 271.9 F 29.6

Min 271.3 F 29.3  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.8 F 0.0  
5:00 243.5 F 0.0  
10:00 195.6 F 0.0  
15:00 181.9 F 0.0  
20:00 197.0 F 0.0  
25:00 215.4 F 0.0  
30:00 207.8 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 24:52  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:03:56  
CYCLE COMPLETE

11/20/2024 12:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_