

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4378

Name: RITTER 1  
Sterilizer ID: V2178198  
03/12/2024 02:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.0 F 0.0  
2:00 98.6 F 0.7  
4:00 119.4 F 2.4  
6:00 146.8 F 4.8  
8:00 180.6 F 7.9  
10:00 202.9 F 11.6  
12:00 211.0 F 2.4  
14:00 216.5 F 6.0  
16:00 225.4 F 9.4  
18:00 234.7 F 13.4  
20:00 234.7 F 9.3  
22:00 242.8 F 11.3  
24:00 243.4 F 12.1  
26:00 254.1 F 17.7  
28:00 264.0 F 23.5  
30:00 272.3 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.6 F 29.2  
2:00 271.9 F 28.6  
3:00 271.5 F 28.8  
4:00 272.5 F 29.2

Min 271.2 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 219.9 F 0.0  
10:00 243.9 F 0.0  
15:00 224.3 F 0.0  
20:00 215.1 F 0.0  
25:00 211.2 F 0.0  
30:00 208.6 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 30:16  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:09:02  
CYCLE COMPLETE

03/12/2024 03:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_