
Midmark M11 - v1.0.0
Total Cycles: 5472

Name: RITTER 1
Sterilizer ID: V2178198
04/03/2025 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 74.8 F 0.4
4:00 91.1 F 1.9
6:00 126.3 F 4.5
8:00 154.4 F 7.4
10:00 177.6 F 10.8
12:00 189.3 F 3.3
14:00 209.6 F 4.2
16:00 220.7 F 7.4
18:00 229.4 F 11.2
20:00 227.4 F 6.4
22:00 239.5 F 11.1
24:00 235.1 F 8.4
26:00 248.6 F 14.7
28:00 259.6 F 20.7
30:00 269.8 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 28.9
2:00 271.8 F 28.6
3:00 272.4 F 28.8
4:00 271.9 F 28.7

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 191.9 F 0.0
10:00 179.7 F 0.0
15:00 185.0 F 0.0
20:00 191.7 F 0.0
25:00 198.2 F 0.0
30:00 201.4 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:57
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:09:11
CYCLE COMPLETE

04/03/2025 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____