
Midmark M11 - v1.0.0
Total Cycles: 3438

Name: RITTER 1
Sterilizer ID: V2178198
06/23/2023 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.3 F 0.0
2:00 176.7 F 3.2
4:00 193.4 F 7.2
6:00 213.4 F 12.3
8:00 225.2 F 6.9
10:00 238.0 F 12.4
12:00 237.6 F 10.0
14:00 236.0 F 8.3
16:00 250.7 F 15.6
18:00 262.8 F 22.6
20:00 272.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.8
2:00 271.6 F 28.4
3:00 272.5 F 28.9
4:00 271.5 F 28.2

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 227.2 F 0.0
10:00 213.3 F 0.0
15:00 214.0 F 0.0
20:00 231.2 F 0.0
25:00 228.8 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:10
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

06/23/2023 03:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____