
Midmark M11 - v1.0.5
Total Cycles: 141

Name: RITTER2
Sterilizer ID: V2313228
06/24/2021 04:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.2 F 0.0
2:00 147.5 F 2.5
4:00 166.1 F 6.8
6:00 195.4 F 12.5
8:00 212.7 F 7.4
10:00 228.9 F 13.3
12:00 235.0 F 11.0
14:00 236.8 F 9.7
16:00 250.4 F 16.4
18:00 262.0 F 23.4
20:00 273.3 F 30.6

STERILIZING
mm:ss Degrees PSI
0:00 273.1 F 30.1
1:00 272.5 F 30.1
2:00 272.6 F 30.2
3:00 272.7 F 30.4
4:00 272.7 F 30.6

Min 272.2 F 30.0
Max 273.1 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 218.4 F 0.0
10:00 218.1 F 0.0
15:00 220.6 F 0.0
20:00 216.7 F 0.0
25:00 218.3 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 20:09
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

06/24/2021 04:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____