

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001344  
12:59 11/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:48  
STERILIZING 3:54  
136.4 C 224kPa 3:54  
136.0 C 222kPa 4:24  
136.4 C 224kPa 4:54  
136.3 C 220kPa 5:24  
136.4 C 224kPa 5:54  
136.2 C 222kPa 6:24  
136.3 C 223kPa 6:54  
136.2 C 222kPa 7:24  
136.2 C 222kPa 7:24  
VENTING 7:24  
AIR DRYING 8:12  
CYCLE COMPLETE 23:12  
-----

Digital Signature #  
E6C1EE6012A5AA5E  
-----