

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001801  
7:22 26/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 3:01  
STERILIZING 4:05  
136.1 C 220kPa 4:05  
136.3 C 223kPa 4:35  
136.4 C 224kPa 5:05  
136.4 C 223kPa 5:35  
136.4 C 224kPa 6:05  
136.3 C 223kPa 6:35  
136.4 C 224kPa 7:05  
136.4 C 225kPa 7:35  
136.4 C 225kPa 7:35  
VENTING 7:35  
AIR DRYING 8:25  
CYCLE COMPLETE 38:25

-----  
Digital Signature #  
54CB6C32C23D3E2B  
-----