

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005223  
10:42 14/03/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:54  
PRESSURIZING 1:58  
STERILIZING 2:52  
136.4 C 221kPa 2:52  
136.2 C 223kPa 3:22  
136.0 C 222kPa 3:52  
136.4 C 223kPa 4:22  
136.4 C 225kPa 4:52  
136.3 C 223kPa 5:22  
136.4 C 224kPa 5:52  
136.1 C 221kPa 6:22  
136.1 C 221kPa 6:22  
VENTING 6:22  
AIR DRYING 7:10  
CYCLE COMPLETE 14:04

-----  
Digital Signature #  
1638EB0DA8B178CA  
-----